



Contents lists available at ScienceDirect

The Journal of Arthroplasty

journal homepage: www.arthroplastyjournal.org

World Expert Meeting in Arthroplasty 2024

Is There a Difference in Outcomes Between Collared and Noncollared Uncemented Femoral Stems in Primary Total Hip Arthroplasty?



Levent Bayam, MD ^a, Kayahan Karaytug, MD ^{b,*}, Oliver Marin-Pena, MD ^c,
Adam Rana, MD ^d, Emilio Romanini, MD ^e, Eryou Feng, MD ^f,
Antonio M. Asensio, MD ^g, Remzi Tozun, MD ^b

^a Department of Orthopedics and Traumatology, İstanbul Medipol University, İstanbul, Turkey^b Acibadem University Maslak Acibadem Hospital, International Joint Center, İstanbul, Turkey^c Hospital Universitario Infanta Leonor, Madrid, Spain^d Department of Orthopedic Surgery, Maine Medical Center, Portland, Maine^e Centre for Hip and Knee Arthroplasty, Polo Sanitario San Feliciano, Rome, Italy^f Fujian Provincial Clinical Medical Research Center for First Aid and Rehabilitation in Orthopaedic Trauma, Fuzhou, PR China^g Hospital General Universitario Reina Sofía, Murcia, Spain

ARTICLE INFO

Article history:

Received 18 September 2024

Received in revised form

16 October 2024

Accepted 18 October 2024

Available online 23 October 2024

Keywords:

collared

noncollared

uncemented

hip

THA

hip arthroplasty

Is there a difference in outcomes between collared and non-collared uncemented femoral stems in primary total hip arthroplasty?

Response/recommendation: The use of both collared and noncollared uncemented femoral stems result in an excellent clinical outcome in patients undergoing total hip arthroplasty (THA). However, the rate of periprosthetic fractures is lower for the collared stems despite similar survivorship.

Level of evidence: Strong.

Delegates vote: Agree 84.6%, Disagree 9.7%, and Abstain 5.8% (strong consensus).

Rationale

The literature is rich with numerous studies, including many high-volume national registries and some systematic review studies, comparing collared versus noncollared uncemented femoral stems in patients undergoing total hip arthroplasty (THA).

One or more of the authors of this paper have disclosed potential or pertinent conflicts of interest, which may include receipt of payment, either direct or indirect, institutional support, or association with an entity in the biomedical field which may be perceived to have potential conflict of interest with this work. For full disclosure statements refer to <https://doi.org/10.1016/j.arth.2024.10.094>.

* Address correspondence to: Kayahan Karaytug, MD, Acibadem University Maslak Acibadem hospital, International Joint Center, Buyukdere Road. no 40, Maslak, Sariyer, İstanbul 34457.

<https://doi.org/10.1016/j.arth.2024.10.094>

0883-5403/© 2024 Elsevier Inc. All rights are reserved, including those for text and data mining, AI training, and similar technologies.

This systematic review was conducted to extract data related to these two different designs of femoral stem.

The studies collected data on important variables such as subsidences, periprosthetic fractures, overall complications, and the need for revision surgery. In addition, patient-reported outcomes (using Western Ontario and McMaster Universities Osteoarthritis Index, Harris hip score, The University of California-Los Angeles score), radiographic outcomes, and survivorships were evaluated. The influence of surgical approach, age, and other variables were also assessed [1,2].

In general, all studies reviewed demonstrated excellent survivorship for both stem types with similar radiographic and functional outcomes [1,3]. However, there was a significant difference between the two stems with regard to the rate of 90-day periprosthetic fracture, subsidence, and rotational instability, with data

favoring collared uncemented femoral stems [4–10]. Alternatively, subsidence may not be clinically relevant as the revision rate is similar [1] and some articles may have a bias on a particular stem type.

Among many recent studies, seven reported on the patients from national registries that included the American, Australian, German, Norwegian, and the United Kingdom joint registries. The studies were conducted over the last 5 years and included a large number of patients that ranged from 59,518 to 337,647 undergoing THA. The main outcome of these studies was periprosthetic fracture incidence and collared stems [4–10].

There were four systematic reviews with slightly different analyses and conclusions. There were two systematic reviews that demonstrated decreased complication rates with long stems as well as collared ones [3,11]. There were three reviews that showed no difference in revision rates [1,3,11]. A single review found lower revision rates with collared stems [2]. The finding of higher revision rates in one review could be due to the use of smaller stems [9]. A systematic review included only randomized controlled studies and found differences in subsidence rates, but no superiority of collared stem [3] and otherwise similar clinical outcomes [1–3,11].

An explanation was offered regarding the differences in the rate of subsidence and periprosthetic fracture seen between the two stems. The collar, resting on the calcar, is believed to provide a protective effect against subsidence and subsequent fracture [2]. The rotational stability of collared stems is also thought to be higher. However, there are some unresolved issues. There is a theoretical issue with the use of collared stems. The collar, particularly when prominent, is thought to be responsible for iliopsoas tendon irritation and subsequent groin pain. Another issue with the use of collared stems relates to the need to seat the stem on the calcar, which in itself can result in intraoperative periprosthetic fractures, which can be treated during the same surgical procedure without requiring subsequent revision surgeries. The rate of intraoperative fractures was not reported in any of the studies that we reviewed.

In conclusion, it appears that the use of collared stems (at least the one common stem that was evaluated) results in a reduction in the rate of periprosthetic fracture, while noncollared stems provide comparable results in terms of long-term survival and functionality. Thus, surgeons starting to familiarize themselves with the use of a femoral stem at the start of their career may want to consider these findings that may influence their choice of a femoral stem.

CRedit authorship contribution statement

Levent Bayam: Writing – review & editing, Writing – original draft, Supervision, Conceptualization. **Kayahan Karaytug:**

Supervision. **Oliver Marin-Pena:** Writing – review & editing, Writing – original draft, Project administration, Conceptualization. **Adam Rana:** Writing – review & editing, Writing – original draft, Conceptualization. **Emilio Romanini:** Investigation, Conceptualization. **Eryou Feng:** Investigation. **Remzi Tozun:** Methodology, Investigation, Formal analysis.

References

- [1] Giovanoulis V, Kenanidis E, Aïm F, Gamie Z, Marmor S, Potoupnis M, et al. Collared versus collarless hydroxyapatite-coated stems for primary cementless total hip arthroplasty; a systematic review of comparative studies. Is there any difference in survival, functional, and radiographic outcomes? SICOT J 2024;10:8. <https://doi.org/10.1051/sicotj/202400>.
- [2] Ait-Si-Selmi T, Vidalain JP, Ramos-Pascual S, Kuratle T, Saffarini M, Dejour E, et al. Outcomes of total hip arthroplasty using collared versus collarless uncemented HA-coated stems: a meta-analysis. EFORT Open Rev 2024 4;9: 264–75. <https://doi.org/10.1530/EOR-22-0091>.
- [3] Nerys-Figueroa J, Parsa A, Curley A, Charif S, Domb BG, Schinsky MF. Slightly reduced early subsidence with similar outcomes and complications rate in collared stems – a systematic review of randomized clinical trials. J Orthop 2024;50:170–6. <https://doi.org/10.1016/j.jor.2024.01.013>.
- [4] Lamb JN, Baetz J, Messer-Hannemann P, Adekanmbi I, van Duren BH, Redmond A, et al. A calcar collar is protective against early periprosthetic femoral fracture around cementless femoral components in primary total hip arthroplasty: a registry study with biomechanical validation. Bone Joint J 2019;101-B:779–86. <https://doi.org/10.1302/0301-620X.101B7.BJJ-2018-1422.R1>.
- [5] Kelly M, Chen AF, Ryan SP, Working ZM, De A, Mullen K, et al. Femoral component design is associated with the risk of periprosthetic femur fracture after cementless THA in patients 65 Years or older. Clin Orthop Relat Res 2024 1;482:1485–93. <https://doi.org/10.1097/CORR.0000000000002985>.
- [6] Tierney J, Jackman E, Holder C, Wall CJ, Wilson CJ. Increased rates of late periprosthetic fractures in larger hydroxyapatite-coated cementless stems: are collared stems a better alternative for total hip arthroplasty? J Arthroplasty 2024;39:744–9. <https://doi.org/10.1016/j.arth.2023.08.068>.
- [7] Konow T, Baetz J, Melsheimer O, Grimberg A, Morlock M. Factors influencing periprosthetic femoral fracture risk. Bone Joint J 2021;103-B:650–8. <https://doi.org/10.1302/0301-620X.103B4.BJJ-2020-1046.R2>.
- [8] Melbye SM, Haug SCD, Fenstad AM, Furnes O, Gjertsen JE, Hallan G. How does implant survivorship vary with different corail femoral stem variants? Results of 51,212 cases with up to 30 years of follow-up from the Norwegian arthroplasty register. Clin Orthop Relat Res 2021;479:2169–80. <https://doi.org/10.1097/CORR.0000000000001940>.
- [9] Hoskins WT, Bingham RJ, Lorimer M, de Steiger RN. The effect of size for a hydroxyapatite-coated cementless implant on component revision in total hip arthroplasty: an analysis of 41,265 stems. J Arthroplasty 2020;35:1074–8. <https://doi.org/10.1016/j.arth.2019.10.060>.
- [10] Morlock M, Perka C, Melsheimer O, Kirschbaum SM. Influence of the type of stem and its fixation on revision and immediate postoperative mortality in elective total hip arthroplasty. Bone Joint J 2024;106-B(3 Suppl A):130–6. <https://doi.org/10.1302/0301-620X.106B3.BJJ-2023-0820.R2>.
- [11] Panichkul P, Bavonratavech S, Arirachakaran A, Kongtharvonskul J. Comparative outcomes between collared versus collarless and short versus long stem of direct anterior approach total hip arthroplasty: a systematic review and indirect meta-analysis. Eur J Orthop Surg Traumatol 2019;29: 1693–704. <https://doi.org/10.1007/s00590-019-02516-1>.